

Tennessee Immunization Program (TIP)

Temperature Log for Refrigerator – Celsius

DAYS 1 - 15

Monitor temperatures closely!

1. Write your initials below in "Staff Initials," and note the time the temperature was checked.
2. Record temps twice each work day.
3. Record the min/max temps once each work day - preferably in the morning.
4. Put an "X" in the row that corresponds to the refrigerator's temperature.
5. If any out of range temps, see instructions to the right.
6. Keep all VFC documents for 3 years.

Month/Year _____ VFC PIN# _____

Facility Name _____

Take action if temp is out of range – (below 2°C or above 8°C)

1. Label vaccine "do not use," and store it under proper conditions as quickly as possible.
2. Do not discard vaccines unless directed to by TIP and/or vaccine manufacturer(s).
3. Record the out of range temp and the room temp in the "Action" box (which correlates with the date) on the bottom of the log.
4. Notify TIP at **1-800-404-3006** immediately for guidance.

Day of Month		1	2	3	4	5	6	7	8	9	10	11	12	13	14	15
Staff Initials																
Exact Time Read		am pm														
Min/MaxTemp (since previous reading)																
Danger! Temperatures above 8°C are too warm! Write any out-of-range temps and room temp on the lines below and call TIP immediately! 1-800-404-3006																
temperatures	8°C															
	7°C															
	6°C															
	Aim for	5°C														
acceptable	4°C															
	3°C															
	2°C															
Danger! Temperatures below 2°C are too cold! Write any out-of-range temps and room temp on the lines below and call TIP immediately! 1-800-404-3006																
action	Out-of-range temps (above 8°C or below 2°C)															
	Room Temperature															

