

October 26, 2020

Robert Colby, Director Chattanooga-Hamilton County Air Pollution Control 6215 Preservation Drive Chattanooga, TN 37416-3638

Reference: Third Quarter 2020 TDEC /Performance Audits

Dear Mr. Colby:

This report contains the results of performance audits conducted by personnel from Tennessee's Air Quality Assurance Program on September 16<sup>th</sup> through 18<sup>th</sup>, 2020. The audits were performed on various instruments at multiple monitoring locations throughout the Hamilton County monitoring network, including gaseous analyzers for ozone and low volume particle samplers for PM 2.5. Sites audited include East Ridge, Eastside Utilities, Siskin Drive and Soddy Daisy.

During the Quality Assurance Performance Audits (QAPAs) on September 16<sup>th</sup> through 18<sup>th</sup>, Tennessee Air Quality Assurance personnel audited the flow rate of each low volume particle sampler against a specified design flow rate and a transfer standard. QA personnel also audited each ozone analyzer with a level three standard (Teledyne 750 multi-gas calibrator) to determine the accuracy of the analyzer against a certified standard. These audits were conducted to ensure that each air monitor within the Hamilton County network was operating within accepted criteria established by federal regulations and guidance documents. Acceptable limits of these performance audits are published in 40 Code of Federal Regulations (CFR) Part 58 and QA Handbook Volume II, which establish federal standards for Quality Assurance and Quality Control practices for ambient air monitoring. These acceptable limits should be considered in conjunction with your Quality Assurance Plan.

Results from the third quarter, 2020 QAPAs, demonstrate that the Hamilton County air monitoring network was operating within acceptable limits of federal regulations and quality assurance guidelines. All QAPA results were noted in the instrument log books and summarized in the following pages of this report.

If we can be of any further assistance or should there be questions regarding this audit, please email me at Ben.Hartsfield@tn.gov or call (615) 532-0597.

Sincerely,

Ben Hartsfield Environmental Consultant Quality Assurance Section

Reviewed by Michelle Oakes Michelle Oakes

cc: A, Pratt, APC; M. Oakes, APC; K. Jones, Hamilton County.



Site:Soddy DaisyAudit Quarter:3Q 2020AQS #:47-065-1011Pollutant:O3Auditor:Ben Hartsfield

Instrument Model	Thermo 49C	<b>O</b> <sub>3</sub>		Audit	Serial #
				Equipment	
Serial Number	58172-316			Teledyne 750U	#71
Data Logger Model	ESC 8872			Teledyne 751H	#183
Data Logger Serial Number	799			Streamline Pro	#150306
Audit Date	9/17/2020			Time	www.time.gov
Ozone Set Point	Audit	Monitor Response	Difference	Acceptable	Audit Level
	Value	(ppb)		Range	
	(ppb)				
Zero	0	0	0		Zero
85 ppb	85	84	1.2%	< <u>+</u> 15.1%	5
65 ppb	65	65	0/0%	< <u>+</u> 15.1%	4
35 ppb	35	35	0.0%	< <u>+</u> 15.1%	3
15 ppb	15	16	6.7%	< <u>+</u> 15.1%	2
Zero	0	0	0		Zero
Data Logger Time	09:56:02	09:56:00	-00:00:02		
Data Logger Temperature	22.5	22.7	0.2		

Remarks:

- 1. The monitor audited showed satisfactory correlation with our audit standard.
- 2. All measured parameters were within acceptable limits.
- 3. The audit results were also noted in the instrument log books.
- 4. The data logger acceptable time range of  $< \pm$  one minute is a State recommendation.\*
- 5. 8816 response 0, 84, 65, 35, 16, 0. 49i response 0, 84, 65, 35, 15, 0.

Recommendations:

1. None.



Site:Eastside UtilityAudit Quarter:3Q 2020AQS #:47-065-4003Pollutant:O3Auditor:Ben Hartsfield

Instrument Model	Thermo 49i	<b>O</b> <sub>3</sub>		Audit	Serial #
				Equipment	
Serial Number	1435663747			Teledyne 750U	#71
Data Logger Model	ESC 8872			Teledyne 751H	#183
Data Logger Serial Number	0642			Streamline Pro	#150306
Audit Date	9/18/2020			Time	www.time.gov
Ozone Set Point	Audit	Monitor Response	Difference	Acceptable	Audit Level
	Value	(ppb)		Range	
	(ppb)				
Zero	0	0	0		Zero
85 ppb	85	85	0.0%	< <u>+</u> 15.1%	5
65 ppb	65	65	0/0%	< <u>+</u> 15.1%	4
35 ppb	35	35	0.0%	< <u>+</u> 15.1%	3
15 ppb	15	15	0.0%	< <u>+</u> 15.1%	2
Zero	0	0	0		Zero
Data Logger Time	10:31:27	10:31:25	-00:00:02		
Data Logger Temperature	21.9	21.8	-0.1		

## Remarks:

- 1. The monitor audited showed satisfactory correlation with our audit standard.
- 2. All measured parameters were within acceptable limits.
- 3. The audit results were noted in the instrument log book and listed above.
- 4. The data logger acceptable time range of  $< \pm$  one minute is a State recommendation.\*
- 5. 8816 response- 0, 85, 65, 35, 15, 0. 49i response- 0, 85, 65, 35, 15, 0.

## **Recommendations:**

1. None.



Site:	Siskin
Audit Quarter:	3Q 2020
AQS #:	47-065-4002
Pollutant:	PM 2.5
Auditor:	Ben Hartsfield

Instrument Model	Thermo 2025i	PM 2.5	Audit Equipment	Serial Number
Serial Number	21084/POC #1		Streamline Pro	#150306
Audit Date	9/16/2020		Time	www.time.gov
Measured Parameter	Audit Value	<b>Monitor Response</b>	Difference	Acceptable Range
<b>Transfer Standard Flow</b>	16.57	16.67	0.6%	< <u>+</u> 4.1%
Design Flow	16.67	16.67	0.0%	< <u>+</u> 5.1%
<b>Barometric Pressure</b>	743.2	742	-1.2	< <u>+</u> 10.0 mm
Ambient Temperature:	27.4	27.1	-0.3	< <u>+</u> 2.1°C
Filter Temperature	28.0	27.8	-0.2	< <u>+</u> 2.1°C
External Leak Check	AF – 5 mmHg	AL – 5 mmHg	Pass	<u>&lt;</u> 25 mmHg/min.
Clock Time	13:35:25	13:35:00	-00:00:25	< <u>+</u> 00:01:00 minute

Instrument Model	Thermo 2025i	PM 2.5	Audit Equipment	Serial Number
Serial Number	21131/POC #2		Streamline Pro	#150306
Audit Date	9/16/2020		Time	www.time.gov
Measured Parameter	Audit Value	Monitor Response	Difference	Acceptable Range
<b>Transfer Standard Flow</b>	16.52	16.67	0.9%	< <u>+</u> 4.1%
Design Flow	16.67	16.67	0.0%	< <u>+</u> 5.1%
<b>Barometric Pressure</b>	742.9	742	-0.9	< <u>+</u> 10.0 mm
Ambient Temperature:	28.2	27.7	-0.5	< <u>+</u> 2.1°C
Filter Temperature	29.0	28.1	-0.9	< <u>+</u> 2.1°C
External Leak Check	AF – 5 mmHg	AL – 5 mmHg	Pass	<u>&lt;</u> 25 mmHg/min.
Clock Time	13:49:26	13:49:00	-00:00:26	< <u>+</u> 00:01:00 minute



Instrument Model	Teledyne T640	PM 2.5	Audit Equipment	Serial Number
Serial Number	83		BGI TriCal	162
Audit Date	9/16/2020		Time	www.time.gov
Measured Parameter	Audit Value	Monitor	Difference	Acceptable Range
		Response		
<b>Transfer Standard Flow</b>	5.02	4.98	-0.8%	$< \pm 5.1\%$
Design Flow	5.0	5.00	0.0%	$\leq \pm 0.15$ LPM
<b>Barometric Pressure</b>	741.0	740.8	-0.2	$< \pm 5.0\%$
Ambient Temperature	26.8	26.2	-0.6	$\leq \pm 2.1^{\circ}\mathrm{C}$
Leak Check	AF: 0.1 lpm	AL: 0.0 lpm	Pass	<±0.25LPM
Data Logger 8832 #A0994K	13:11:46	13:11:47	00:00:01	*< ± 00:01:00 minute
Time:				

Remarks:

- 1. The instruments audited showed satisfactory correlation with our audit standards.
- 2. The PM 10 inlets and PM 2.5 cyclones were clean.
- 3. The audit results were noted in the instrument log books.
- 4. The data logger acceptable time range of  $< \pm$  one minute is a State recommendation. \*

Recommendations:

1. None.



Site:	East Ridge
AQS #:	47-065-0031
Audit Quarter:	3Q 2020
Pollutant:	PM 2.5
Auditor:	Ben Hartsfield

Instrument Model	Thermo 2025	PM 2.5	Audit Equipment	Serial Number
Serial Number	22109		Streamline Pro	#150306
Audit Date	9/17/2020		Time	www.time.gov
Measured Parameter	Audit Value	<b>Monitor Response</b>	Difference	Acceptable Range
<b>Transfer Standard Flow</b>	16.55	16.67	0.7%	< <u>+</u> 4.1%
Design Flow	16.67	16.67	0.0%	< <u>+</u> 5.1%
<b>Barometric Pressure</b>	738	737	-1.0	< <u>+</u> 10.1 mm
Ambient Temperature	27.6	26.8	-0.8	< <u>+</u> 2.1°C
Filter Temperature	28.0	27.1	-0.9	< <u>+</u> 2.1°C
Compartment Temp.	27.6	27.2	-0.4	< <u>+</u> 2.1°C
External Leak Check	AF: 5 mm Hg/min	AL: 4 mm Hg/min	Pass	< 25 mm Hg/min
Clock Time	12:14:56	12:15:00	00:00:04	< <u>+</u> 00:01:00 minute

Remarks:

- 1. The monitor audited showed satisfactory correlation with our audit standard.
- 2. PM 10 inlet and PM 2.5 cyclone were clean.
- 3. The results were noted in the instrument log book.

Recommendations:

1. None.

Any questions concerning the audits or any comments made in connection with the audits may be addressed to me at (615) 532-0597 or Ben.Hartsfield@tn.gov.

Sincerely,

Ben Hartsfield Environmental Consultant Quality Assurance Section

Reviewed by Michelle Oakes Mar Ol Report Date: 10/26/2020

To: R. Colby, Hamilton Co.

cc: K. Jones, Hamilton Co.; A. Pratt, APC; M. Oakes, APC.