

6-8-21 TDEC QAPA Amer. steel High

Vol Pb # P 4304 - QA 39.53

QM 40.96 % 3.6% DLA-Lance A

AE - Orifice 10337 - SLP 060504

Cloudy : Calm - 79°F DLA-Lance A

6-8-21 DC 11:45 EST 28°C P. Cloudy

6-8-21 DC 8:00 EST 23°C Cloudy

-state Audit with Lance Allen

6-8-21 TDEC QAPA High Vol Pb

# P04302 - QA 40.63 QM 40.57 -0.1%

AE - Orifice 10337 - SLP 060504 - (DLA)

Cloudy - light winds - 77°F Lance Allen

6-8-21 DC 7:50 EST 23°C Cloudy

-state Audit with Lance Allen

6-8-21 TDEC QAPA High Vol P02875

QA 40.32 QM 40.04 -0.6% P.P

AE - Orifice 10337 - SLP 060504

Cloudy - light winds - 77°F - DLA-Lance A

6-9-21 TDEC QAPA Springhill 03 - Teledyne T400 #4005-  
 DL 8832 #A375BK / AE Teledyne 750 #112-751H #182-SLP060504  
 QA zero 85 65 35 15 zero 7:49:06 23.7  
 QM 0.7 85 65 36 15 110 7:49:06 23.7  
 PT Cloudy : Calm - (DLA) - Lance Allen

6-9-21 TDEC QAPA East Knox 03 - Teledyne T400 #4006-  
 DL 8832 #A375TK / AE - Teledyne 750 #112-751H #182-SLP  
 060504  
 QA 9:56:00 24.6 zero 85 65 35 15 zero  
 QM 9:56:00 24.1 -0.7 83 63 33 14 -0.9  
 PT Cloudy : Calm - (DLA) - Lance Allen

6-8-21 TDEC QAPA Air Lab 2025 #22576  
 QA 31.7 32.5 30.0 735 16.81 12:29:27 AF=3  
 QM 31.8 32.2 29.8 736 16.70 12:29:00 AL=2  
 1 CompT FilT ambT ambP Flow Time  
 AE-SLP 060504 - (DLA) - Lance Allen

6-9-21 TDEC QAPA Rule Thermo 2025:  
 #20952 / AE - SLP060504 / Leak ch  
 QA 12:55:13 27.8 26.3 731 16.88 AF=5  
 QM 12:55:00 27.1 26.1 736 16.67 AL=4  
 Time FilT ambT ambP Flow  
 DLA  
 6-9-21 PT Cloudy : Light Winds  
 (DLA) - Lance Allen  
 AT - E 11 C 1

6-8-21 TDEC QAPA Springhill Sass

# 69188-69162/AE-SLP 060504

	Ch1	Ch2	ambT	ambP	Ch1	Ch2	Time
QA	26.3	26.3	24.8	737	6.79	6.65	9:44:23

QM	25.8	26.2	25.3	734	6.7	6.7	9:42:00
----	------	------	------	-----	-----	-----	---------

Leak ch 1=0.0 2=0.0 Pt Cloudy: Calm (DLA)

6-8-21 Springhill URG # B0409-B0224

AE-TriCal #162

QA	26.8	736.5	22.88	10:11:19
----	------	-------	-------	----------

QM	26.8	736.1	22.02	10:13:00
----	------	-------	-------	----------

Pt Cloudy: Calm - DLA - Lance Allen

6-9-21 TDEC QAPA AirLab Teledyne TG40x

#192/AE SLP 060504/Leakch AF=0.0/0.0

QA	16.82	5.08	11.66	735	26.8	13:43:00
----	-------	------	-------	-----	------	----------

QM	16.70	5.01	11.67	734	26.8	13:43:00
----	-------	------	-------	-----	------	----------

Leakch AL=0.0/0.0 / Cloudy: Windy - (DLA)

Lance Allen

6-9-21 TDEC QAPA Rule Teledyne - T640#675

QA 13:13:00 5.09 26.5 731 Leak ch AL

QM 13:13:00 5.01 26.7 732 AF 0.0/0.0 0.0/0.0

AE - SLP 060504 - Cloudy! Light wind (DLA)

6-8-21 TDEC QAPA Springhill Teledyne

T640 #910 / AE SLP 060504 / Leak ch

QA 5.12 29.7 734 10:21:00 0.0/0.0

QM 5.00 28.2 737 10:21:00 0.0/0.0

P+Cloudy! Calm - DLA - Lance Allen